



## DECLARATION OF CONFORMITY

*Apparatus Identification:* **GigatestPro**

*Apparatus Classification:* Test and Measurement Equipment

*Statement of Conformity:*

Based on sample product test result using appropriate standards, and in accordance with the following EC directives ILLKO, s.r.o. hereby declares the GigatestPro to be in conformity with:

EC Directive 89/336/EEC amended by 91/263/EEG, 92/31/EEC

EC Directive 72/23/EEC amended by 93/68/EEC

**Sample Product Testing for EMC:**

*Tested by:* INSTITUTE FOR TESTING AND CERTIFICATION, a.s.  
Testing Laboratory No. 1004.3  
Division 4 - MESIT QM  
Sokolovská 573  
Uherské Hradiště  
Czech Republic

*Standards Used:*

ČSN EN 55022 ed.2: 2007	(Czech version EN 55022 : 2006)
ČSN EN 61326-1 : 2006	(Czech version EN 61326-1 : 2006)
ČSN EN 61000-4-2 : 1997 + A1 : 1999	(Czech version EN 61000-4-2 : 1995 + A1 : 1998)
ČSN EN 61000-4-3 ed.3 : 2006	(Czech version EN 61000-4-3 : 2006)
ČSN EN 61000-4-4 ed.2: 2005	(Czech version EN 61000-4-4 : 2004)
ČSN EN 61000-4-5 ed.2 : 2007	(Czech version EN 61000-4-5 : 2006)
ČSN EN 61000-4-6 : 1997 + Z : 2001	(Czech version EN 61000-4-6 : 1996 + A1 : 2001)

*Report ID:* Test Report No. 3681/09 – Test Report about Electromagnetic Test on the Measuring Instrument IL1700 ISO

**Sample Product Testing for safety:**

*Tested by:* INSTITUTE FOR TESTING AND CERTIFICATION, a.s.  
Testing Laboratory No. 1004.3  
Division 4 - MESIT QM  
Sokolovská 573  
Uherské Hradiště  
Czech Republic

*Standards Used:*

ČSN EN 61010-1 : 2003	(Czech version EN 61010-1 : 2001)
ČSN EN 61010-31 : 2003 + A1 : 2008	(Czech version EN 61010-31 : 2002)

*Report ID:* Test Report No. 3339/09 – Test Report about Electric Safety Test on the Measuring Instrument IL1700 ISO

ILLKO s.r.o.  
MASARYKOVA 2226  
678 01 BLANSKO

*Ivo Lipovský*  
Ivo Lipovský  
Product Manager

Issue Date: 1.11.2009